

Verification of Compliance

Product Name : TravelScope Series Digital Storage Oscilloscope
Brand Name : [Acufe](#)
Model Number : TS2212, TS2202
Applicant : Acute Technology Inc.
Address : 6F-7, #12, Ln. 609, Sec. 5, Chongxin Rd., Sanchong Dist., New Taipei City 24159, Taiwan
Report Number : C-A520-1210-322
Issue Date : November 9, 2012
Applicable Standards : EN 61326-1:2006 Class B
EN 55011:2009+A1:2010 Group 1 Class B
EN 61000-3-2:2006+A1:2009+A2:2009
EN 61000-3-3:2008
EN 61326-1:2006
EN 61000-4-2:2009
EN 61000-4-3:2006+A1:2008+A2:2010
EN 61000-4-4:2004+A1:2010
EN 61000-4-5:2006
EN 61000-4-6:2009
EN 61000-4-11:2004

Based on the EMC Directive 2004/108/EC and the specifications of the customer, one sample of the designated product has been tested in our laboratory and found to be in compliance with the EMC standards cited above.



TAF 0905
FCC CAB Code TW1053
NVLAP Lab Code 200575-0
IC Code 4699A
VCCI Accept. No. R-1527, C-1609, T-1441, G-10



Central Research Technology Co.
EMC Test Laboratory
11, Lane 41, Fushuen St., Jungshan Chiu,
Taipei, Taiwan, 104, R.O.C.
Tel : 886-2-25984568
Fax: 886-2-25984546

(Tsun-Yu Shih/ General Manager)

Date: November 9, 2012

Verification of Compliance

Product Name : TravelScope Series Digital Storage Oscilloscope
Brand Name : [Acufe](#)
Model Number : TS2212, TS2202
Applicant : Acute Technology Inc.
Address : 6F-7, #12, Ln. 609, Sec. 5, Chongxin Rd., Sanchong Dist., New Taipei City 24159, Taiwan
Report Number : F-A520-1210-322
Issue Date : November 9, 2012

Applicable Standards : FCC Part 15, Subpart B Class B
ANSI C63.4:2009

One sample of the designated product has been tested in our laboratory and found to be in compliance with the FCC rules cited above.



NVLAP LAB CODE 200575-0

TAF 0905

FCC CAB Code TW1053

IC Code 4699A

VCCI Accep. No. R-1527, C-1609, T-1441, G-10




Central Research Technology Co.

EMC Test Laboratory

11, Lane 41, Fushuen St., Jungshan Chiu,
Taipei, Taiwan, 104, R.O.C.

Tel : 886-2-25984568

Fax: 886-2-25984546



(Tsun-Yu Shih/ General Manager)

Date: November 9, 2012



VERIFICATION

of conformity with Low Voltage Directive

Verification No.: ACT202615LVD

Document holder:

Acute Technology Inc.

Address:

6F-7, #12, Ln. 609, Sec. 5, Chongxin Rd.,
Sanhong Dist., New Taipei City 24159, Taiwan

Trade mark:

Acute.
PC-based T&M Instruments

Type of product:

TravelScope Series Digital Storage Oscilloscope

Type designation:

TS2212, TS2202

Technical data:

5Vdc, USB 2.0, CAT I, Class III

A sample of the product has been assessed with respect to CE-marking according to the Low Voltage Directive (2006/95/EC) and found to comply with the essential requirements of the Directive. The Standard(s) used for showing the compliance and the full details of the results are given in the Test Report as detailed below:

Standard(s)	Report No.	Report Issued Date
EN 61010-1:2010 (Third Edition) and EN 61010-2-030:2010 (First Edition)	ACT202615	October 30, 2012

The holder of the verification is authorized to use this verification in connection with the EC declaration of conformity according to the Directive. The CE marking may only be used if all relevant and effective EC Directives are complied with. Together with the manufacturer's own documented production control, the manufacturer (or his European authorized representative) can in his EC Declaration of Conformity verify compliance with the Low Voltage Directive.

Approved by:

Vincent Tan

Acts Certification and Testing Services

October 30, 2012



Acts Certification and Testing Services Co., Ltd.

1st Floor, No. 28, Lane 61, Tianshang Road, Taipei, Taiwan Tel: +886-2-25858775 Fax: +886-2-25983002





VERIFICATION

of conformity with CE Directives

Verification No.: ACT202615CE

Document holder:
Acute Technology Inc.

Address:
6F-7, #12, Ln. 609, Sec. 5, Chongxin Rd., Sanzhong
Dist., New Taipei City 24159, Taiwan

Trade mark:

Acute.
PC-based T&M Instruments

Type of product:
TravelScope Series Digital Storage
Oscilloscope

Type designation:
TS2212, TS2202

Technical data:
5Vdc, USB 2.0, CAT I, Class III

A sample of the product has been assessed with respect to CE-marking according to the Low Voltage Directive (2006/95/EC) and Electromagnetic Compatibility Directive (2004/108/EC) and found to comply with the essential requirements of the Directives.

The Standard(s) used for showing the compliance and the full details of the results are given in the Test Reports as detailed below:

Standard(s)	Report No.	Report Issued Date
EN 61010-1:2010, EN 61010-2-030:2010	ACT202615	October 30, 2012
EN 61326-1:2006, EN 55011:2009+A1:2010, EN 61000-3-2:2006+A1:2009+A2:2009, EN 61000-3-3:2008, EN 61000-4-2:2009, EN 61000-4-3:2006+A1:2008+A2:2010, EN 61000-4-4:2004+A1:2010, EN 61000-4-5:2006, EN 61000-4-6:2009 and EN 61000-4-11:2004	C-A520-1210-322	November 9, 2012

The holder of the verification is authorized to use this verification in connection with the EC declaration of conformity according to the Directives. The CE marking may only be used if all relevant and effective EC Directives are complied with. Together with the manufacturer's own documented production control, the manufacturer (or his European authorized representative) can in his EC Declaration of Conformity verify compliance with the Directives.

Approved by:

Vincent Tan
Acts Certification and Testing Services
November 9, 2012



Acts Certification and Testing Services Co., Ltd.

1st Floor, No. 20, Lane 61, Tianshiang Road, Taipei, Taiwan Tel: +886-2-25858775 Fax: +886-2-25983802

